

What Really Happened?

Treatment Options and Decision

- Options

- Surveillance – AAA sac, 5.5cm, no growth demonstrated
- Graft Explant – Supraceliac clamp, high surgical risk
- FEVAR – exclude right renal stent
- Relining with AFX Bif + Vela Cuff + SMA Chimney
- Relining with Vela Cuff + Vela Cuff + SMA Chimney

Rationale for Treatment Approach

- Medical History
- Patient Preference
- Original Device Design Characteristics
- Urgency of intervention
- Interaction Between Re-intervention and Original Device

Treatment

Attempted/Achieved

Type IIIb

Relined with Palmaz → Type IIIb persisted

Relined with Vela Cuff → Type IIIb resolved

Type Ia

SMA Chimney + Vela Cuff

Seal extended to CA



Introduce the “unblinded”
engineer speaker